



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N 130144 ✓
Name: Jose Ramirez	ID# 6532	Date 15 October 2008 ✓
A Agency 3 rd Security Forces Squadron		Phone # (907)552-6570
Instrument Location: Law Enforcement desk, Elmendorf AFB, AK ✓		
B Alco S/N X301755 ✓ Target Value .079 ✓ High Pressure 875psi		
Alco Test Values 0.080 0.079		
1 st Alco 2 nd Alco		
Signature 		 11/20/08
(OVER)		


Under the Alaska Rules of Evidence, I certify that, (Do Not write in the area below)

(a) I am Jeanne Swartz, Forensic Scientist III, Scientific Crime Detection Laboratory, State of Alaska Department of Public Safety, and was, at the time of signing, the custodian of the records of the aforementioned laboratory and the signature was made in my official capacity as Scientific Director Designee of the State Breath Alcohol Testing Program.

(b) The verification of calibration for the above instrument has been received and reviewed. Instrument operator, Jose A Ramirez, whose name appears in my records, is qualified to verify the calibration of the above instrument.

(c) This instrument is certified for evidentiary use and these records are being retained as original documents in my custody.

Signed and sealed



Jeanne Swartz
Scientific Director Designee
State of Alaska
State Breath Alcohol Testing Program

Date

11/21/08

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130144 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

OCTOBER 15, 2008 ✓ *CAB 11/20/08*

OPERATOR'S NAME:

RAMIREZ

OPERATOR'S NUMBER: 6532

SUBJECT'S LAST NAME:

VOC

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: 1234567890

DEPT/AGENCY: EDF4

CASE/REPORT: 20081015

TEST TYPE: U

ALCO TARGET VALUE: .079 ✓

ALCO S/N: X301755

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 29.59 in
ALCO TARGET .078 12:53
BLANK TEST .000 12:54
INTERNAL STANDARD VERIFIED 12:54
ALCO TV 29.59 in .000 ✓ 12:54
BLANK TEST .000 12:55
SUBJECT SAMPLE *CAB 11/20/08* .000 12:56
BLANK TEST .000 12:56
ALCO TV 29.59 in .079 ✓ 12:57
BLANK TEST .000 12:57

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130144

OCTOBER 15, 2008 ✓
TIME 12:58

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 50c

BREATH TUBE: 42c

BAROMETER: 29.59 in

FLOW DETECTOR: OKAY

PUMP
HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
GHIJKLMNOPQRSTUVWXYZ[\]^_`abcde fghij klmno
pqrstuvwxyz{|}~